

Truck Scale Report & Certification



Cell☎ (206) 375-1455
 E-mail✉ raysscalederek@yahoo.com
 PO Box 8282
 Bonney Lake, WA 98391

Date Tested: April 25, 2025

Company Name and Address: Ellensburg Cement Products
 Clerf Pit
 930 Prater Road

Balance on Arrival:	0	Min Grad:	20 lbs	Size Platform:	80' x 10'
Capacity:	60 ton	CLC:	30 ton	Scale:	Murphy
Make:	Rice Lake	Serial #:	151935	County:	Kittitas
Number of Sections:	4	Scale Serial #:	N/A	Reg Element:	IQ 355

Loaded Postion	Test Weights	Scale Indication	ERROR	Load Position	Test Weights	Scale Indications	ERROR
As Found				As Left			
Section 1	Test Truck	32,640	0	Section 1			
Section 2	Test Truck	32,640	0	Section 2			
Section 3	Test Truck	32,620	0	Section 3			
Section 4	Test Truck	32,640	0	Section 4			
Section 5	Test Truck	32,620	0	Section 5			
Section 6	Test Truck	32,660	0	Section 6			
Section 1	12,000	12,000	0	Section 1			

Strain Load Test			
Empty Truck Weight	Test Weights	Total	Error
20,640	12,000	32,640	0

Condition upon Inspection							
Stands	<i>Good</i>	Deck	Good	Indicator	Excellent	Cables	Fair
Cell	<i>Good</i>	Bolts	Fair	Printer	N/A	J-Boxes	Poor
Links	<i>Fair</i>	Guide Rails	Good	Scoreboard	N/A	Coping	Good

Remarks and / or Instructions

Scale Technician (Signature and Agent #): *Derek Larsen* # 16705 Owner / Operator: